

Backstop in the right position during injection

EN INSTRUCTIONS FOR USE FOR BELOTERO® REVIVE

1 Description

BELOTERO Revive is a sterile, non-pyrogenic, viscoelastic, colourless, transparent cross-linked sodium hyaluronate gel of non-animal origin in a physiological phosphate buffer containing glycerol.

2 Presentation

BELOTERO Revive is presented in a single use pre-filled glass syringe sterilized by moist heat. Each box contains one instruction leaflet, one or several syringe(s), two traceability labels per syringe and sterile CE-marked needles for single use only. The number of syringes, the dimensions and the number of needles are stated on the external box.

3 Composition

Cross-linked sodium hyaluronate: 20 mg/ml

Glycerol: 17.5 mg/ml

Phosphate buffer pH 7 q.s.: 1.0 ml

Sodium hyaluronate is produced by fermentation of *Streptococcus equi*.

4 Intended Use/Indications

4.1 Intended Use

BELOTERO Revive is an injectable resorbable implant intended to treat early-signs of photodamaged skin via rehydration of dry and very dry skin, and smoothening of superficial fine lines.

4.2 Indications

BELOTERO Revive is indicated for treatment of early signs of photodamaged facial skin, as characterized by dehydration and presence of superficial fine lines.

5 Posology and administration method

BELOTERO Revive is designed to be injected into mid- to deep dermis by authorized health care professionals in accordance with local applicable regulation who have appropriate training, experience, and who are knowledgeable about the anatomy at and around the site of injection in order to minimize the risk of potential complications.

Inject BELOTERO Revive slowly and not too fast to apply the least amount of pressure necessary, according to the appropriate injection technique using the provided needles.

The risk of an intravascular injection can be reduced by different strategies, including aspiration prior to injection, utilizing lower volumes and serial injections in high-risk areas, treating one side at a time, pinching/tenting the skin to provide more space superficial to the branches of the main arteries, and manual occlusion of the origin of the supratrochlear vessels with the non-dominant finger. Blunt cannulas may reduce, but not eliminate the risk.

General recommended injection techniques are for example: linear or serial threading, fanning, cross-hatching or serial (micro)puncture. The quantity of product to be injected depends on the area to be corrected.

BELOTERO Revive must be injected under appropriate aseptic conditions into healthy, non-inflamed skin. Before injection, thoroughly disinfect the area to be treated.

To ensure optimal use of BELOTERO Revive, it is recommended to assemble the needle according to the diagrams below. Improper assembly may lead to a separation of the needle and syringe and / or leakage of the gel at the Luer-lock connection during injection and may cause injury to the patient and/or physician. If during the injection, needle disengagement or a gel leak occurs, discard the syringe and the needle and restart the procedure with a new product.

If the needle becomes obstructed and the injection pressure is too high, stop the injection and change the needle. Never try to straighten a bent needle; instead throw it away and replace it.

The quantity of the gel to be injected depends on the area to be treated and the correction to be achieved. Do not over-correct.

Belotero Revive can be injected under a treatment regimen of 1 to 3 injection sessions. The shortest possible reinjection interval is 4 weeks. This treatment regimen can be renewed at earliest after 6 months, resulting in a maximum annual dose of 24 mL.

The graduations on the syringe label are only intended for orientation for the user.

Gently massage the treated area after the injection to distribute the product uniformly.

Before treatment, the patient's suitability for the treatment and the patient's need for pain relief (topical anaesthetics, ice packs, distraction techniques, local anaesthetic injections, or nerve blocks depending on the injection site and size of needle used), should be assessed.

6 Contra-indications

BELOTERO Revive is contra-indicated:

- In case of known hypersensitivity to one of the product's components, especially to sodium hyaluronate or to glycerol or BDDE,
- In pregnant and breast-feeding women,
- In patients under 18 years old,
- In patients presenting a general infection,
- In patients presenting an active auto-immune disease.

Do not inject BELOTERO Revive into blood vessels.

Do not inject BELOTERO Revive into skin areas presenting active cutaneous inflammation or infection due to e.g. immunological, allergic, bacterial, fungal or viral causes.

Do not inject BELOTERO Revive into an area previously treated with a permanent dermal filler.

Do not inject BELOTERO Revive into the glabellar or nose region.

7 Precautions for use

Health care practitioners are encouraged to discuss all potential risks of soft tissue injection with their patients prior to treatment and ensure that patients are aware of signs and symptoms of potential complications.

In the absence of available clinical data on tolerance of the injection of BELOTERO Revive in patients presenting a history of severe multiple allergies or anaphylactic shock, the practitioner must decide whether to inject BELOTERO Revive on a case-by-case basis depending on the nature of the disease as well as the associated treatment as it may worsen the existing patient health condition. It is recommended to propose a prior double test to these patients and not to inject if the disease is evolving. It is also recommended to carefully monitor these patients after injection.

BELOTERO Revive can be used in combination with other Belotero® products during the same session but in different facial areas. Instructions for use of each product should be followed.

No clinical data is available on the injection of BELOTERO Revive into patient with a Fitzpatrick skin type V/VI.

No clinical data are available on the injection of BELOTERO Revive into the hands.

It is recommended not to inject BELOTERO Revive in patients with a history of streptococcal diseases and in patients predisposed to hypertrophic scars or keloids.

BELOTERO Revive can be used in combination treatments such as with botulinum toxin and/or calcium hydroxylapatite (Radiesse®) only if injected in different facial areas. Practitioners should be experienced and patients appropriately selected as not only benefits but also adverse events can be cumulative and causality of adverse events could become difficult to determine. Instructions for use, depth of injection and appropriate recommendation of each product should be followed. No clinical data are available on the injection of BELOTERO Revive into an area already treated with other aesthetic products or procedures.

BELOTERO Revive must not be used in association with other aesthetic techniques such as peeling, dermabrasion, or any type of laser treatment before complete healing of the last treatment. In any case, even if the healing occurs earlier, BELOTERO Revive must not be used earlier than 2 weeks after the last treatment. No clinical data is available on the combined use of BELOTERO Revive with the above-mentioned treatments.

Patients using anti-coagulation, anti-platelet, or thrombolytic medications (e.g. warfarin), anti-inflammatory drugs (oral/injectable corticosteroids or non-steroidal anti-inflammatory drugs (NSAIDs; e.g. aspirin, ibuprofen)), or other substances known to increase coagulation time (vitamins or herbal supplements, e.g. Vitamin E, garlic, Ginkgo biloba and St. John's Wort), from 10 days pre- to 3 days post-injection may have increased reactions of hematomas, nodules or bleeding at the injection site. Injection of BELOTERO Revive into patients with a history of previous herpetic eruption may be associated with herpes reactivation and HHV related diseases (e.g. pityriasis rosea).

In cases of patients suffering from epilepsy, impaired cardiac conditions, severely impaired hepatic function or severe renal dysfunction or porphyria, the practitioner must decide whether to inject BELOTERO Revive on a case-by-case basis depending of the nature of the disease as well as the associated treatment.

Check the integrity of the inner packaging and the expiry date for both the syringe and the needle prior to use. Do not use these products if the expiry date has lapsed or if the inner packaging has been opened or damaged.

Do not transfer BELOTERO Revive into another container and do not add other substances to the product.

Only the gel is sterile, but not the outside of the syringe.

Discard the syringe, the remaining product and the needles in the appropriate container after use.

Do not re-sterilize and do not reuse due to the associated risks including infection.

The patient must avoid applying makeup (including skin care products) for at least 12 hours after treatment as well as avoid saunas, peeling, Turkish baths and prolonged exposure to the sun, UV rays, extreme heat and cold for 2 weeks after the treatment. Patients should also avoid putting pressure on and/ or handling the treated area and should avoid strenuous physical activity following treatment.

The patient must avoid drinking alcohol for 24 hours before and after treatment. Alcohol may cause the blood vessels to dilate and cause more bruising.

8 Warnings

- Sodium hyaluronate precipitates in the presence of quaternary ammonium salts (such as benzalkonium chloride). It is therefore recommended that BELOTERO Revive does not come into contact with such substances.

- Rare but serious adverse events associated with the intravascular injection of soft tissue fillers in the face have been reported and include temporary or permanent vascular complication, vision impairment, blindness, cerebral ischemia or cerebral hemorrhage, leading to stroke, skin necrosis, and damage to underlying facial structures. Practitioners should immediately stop the injection if a patient exhibits any of the following symptoms, including changes in vision, signs of a stroke, blanching of the skin, or unusual pain during or shortly after the procedure. Patients should receive prompt medical attention and possibly evaluation by an appropriate health care practitioner specialist should an intravascular injection occur.

9 Side effects and adverse events

Patients must be informed by the practitioner about possible side effects and adverse events before treatment.

- Side effects:

Injection site reactions may occur following injection into the skin but disappear spontaneously within a few days. This includes swelling, nodule or lump/bump, bruising/purpura, hematoma, ecchymosis, induration, erythema/redness, tenderness, pain, discoloration and pruritus/itching, tingling, paraesthesia, numbness, hypoaesthesia, scabbing, needle mark and discomfort or irritation. These injection site reactions are generally of mild or moderate intensity. A transient bleeding may also occur at the injection site and usually stops spontaneously as soon as the injection is finished.

- Adverse events:

In occasional cases, one or more of the following may occur in conjunction with the use of products of the Belotero portfolio either immediately or as a delayed reaction: acne cystic, milia, skin dryness (rough facial skin, exfoliation), injection site erosion, inflammation, shivering, fatigue, lymphatic system disorder, rash, burning sensation, injection site sore/warmth/fever, pruritus/itching, urticaria, hematoma, telangiectasia, ecchymosis, edema (including lymph edema), headache/cephalgia, tumefaction, tension, swelling (including persistent swelling), hyper- or hypo-pigmentation, angioedema, induration, blister, vesicle, papule, lump/ bump (visible and/or palpable material) or nodule (including inflammatory nodules), mass, granuloma (including inflammatory signs and foreign body reactions), necrosis, ischemia, vascular occlusion, embolization, infarction, Tyndall effect (including translucent chords), hypersensitivity, allergic reactions (including asthma attack, Quincke's edema, anaphylactic shock or throat tightening) to one of the product's components (e.g. hyaluronic acid, BDDE), oral and dental disorders, nervous system impairment, impairment of the otorhinolaryngological system (e.g. nasal congestion, oropharyngeal pain, dysguesia, rhinorrhea, epistaxis, sinusitis, transient hearing loss), mastication pain, parotid gland enlargement, muscle twitching, muscle injury/disorder, nausea, vomiting, circulatory collapse, presyncope, peripheral venous disease, hot flush, anxiety caused by trypanophobia, patient dissatisfaction and disappointment (due to lack of or reduced performance, decreased firmness/response, undesirable aesthetic effect), injection site discharge, device migration, product distribution issue (e.g. product accumulation), injection site indentation, superficial vein prominence, overcorrection or cranial nerve disorder (e.g. cranial nerve paralysis, facial paralysis, trigeminal neuralgia).

Rare cases of the following adverse events have been reported with hyaluronic acid products such as infection (e.g. erysipelas, phlegmon, cellulitis, including open or draining wounds and (dental) abscess, impetigo, pustules), chronic infection (including biofilm formation), scarring, persistent skin discoloration, sensory dysfunction, non-thrombotic lung embolism as well as sarcoid granuloma formation in subjects with hepatitis C and interferon treatment, cerebral injuries (e.g. intracranial penetration, subarachnoidal hemorrhage), strabismus, ophthalmoplegia, iris adhesions, cataract, conjunctival hemorrhage, eyelid ptosis and lacrimation.

The risk of granuloma, ischemia, necrosis and vascular occlusion is higher with deep injections and high volumes.

Isolated cases of visual impairment or blindness following unintentional intra-arterial injection have been reported in literature.

Patients should be instructed to report any side effects which last for more than one week and any adverse event as soon as it occurs to his/her practitioner, especially if patient has changes in his/her vision, signs of a stroke (including sudden difficulty speaking, numbness or weakness in his/her face, arms, or legs, difficulty walking, face drooping, severe headache, dizziness, or confusion), white appearance of the skin, or unusual pain during or shortly after treatment. The practitioner may then refer the patient to the appropriate treatment.

Patients with lighter skin types are more likely to develop injection-related adverse events. However, patients with skin of color are more likely to develop post-inflammatory hyperpigmentation and /or hypertrophic scar/keloid formation following injection procedures. Patients with specific ethnic characteristics, e.g. Asian population, should be informed of a higher risk of tissue reactions, e.g. itching, swelling, erythema, inflammation.

There is no known interaction with local or loco-regional anesthetics.

10 Assembly of needle to syringe

For optimal use of BELOTERO Revive it is important that the needle is properly connected to the syringe. See diagrams 1, 2, 3 and 4.

1. **Firmly hold** the glass cylinder of **the syringe and the Luer-lock adaptor** between the thumb and the index fingers.
2. Grasp the protective cap with the other hand and unscrew it.
3. **Push & Twist** the needle on the syringe **until a resistance is felt**. Do not over-tight. Over-tightening of the needle may lead to the Luer-lock moving and dislodging from the syringe.
4. Keep holding the Luer-lock and remove the sheath from the needle.

11 Storage







Store between 2 °C and 25 °C. Protect from light and freezing. Avoid mechanical shocks.

12 References

Updated documentation may be available from ANTEIS SA, Switzerland.

13 Instructions how to use the patient implant card

An implant card is provided with BELOTERO Revive. This implant card must be completed by the physician according to the below instructions and provided to the patient after injection.

Number	Symbols	Details
1		Please enter the person's name
2		Please enter the date of implantation
3 - 4		Please enter the name and address of healthcare professional
5		Please enter the number of injections
6		Please enter the total volume injected
7		Please enter the injection site(s)
8	Product label	Please stick here one of the product traceability label, you can keep the second one for your records.

Important information to be provided to the patient

Instruct the patient to keep the patient implant card with her/him and to present it to her/his healthcare professional in case of other appointments. Information about previous treatment must be presented to her/his healthcare professional before treatment!



Caution



Single use product. Do not re-use



Temperature limit of storage: 2 °C – 25 °C



Use-by date



Date of manufacture



Sterile. Sterilized by moist heat. Only the gel is sterile, but not the outside of the syringe.



Consult instructions for use (electronic instruction for use)



Do not use if package is damaged



Open the blister by pulling the Tyvek lid following the arrow



Batch code



CE mark in accordance with Directive 93/42/EEC or Regulation (EU) 2017/745 relating to medical devices. This mark is followed by the notified body number.



Manufacturer



Sterile. Sterilized by irradiation. Only the needle itself is sterile, but not the outside of the needle packaging.

Manufacturer of BELOTERO Revive

Anteis SA

18 Chemin des Aulx
1228 Plan-les-Ouates
Switzerland

Australian Sponsor Name and Address:

Merz Australia Pty Ltd

Suite 8.01A,
189 O'Riordan Street
Mascot, NSW, 2020
Australia

For further information or to request a paper version of the IFU, please call: Tel: +61 2 8076 8139

Date of the instructions for use

2023-09-22